

Product Name	7G One® EnzMix		
Cat NO	7G1-E81	QC Date	Nov 21, 2025
Lot NO	B030310	Report Date	Nov 21, 2025
QC Items	End products	Store at	2°C to 8°C for 2 years
According to	Enterprise standard ZS-B-ZLBZ-TY document	Expire Date	Nov 20, 2027
QC Standard	According to the SOP process document developed by ISO13485: 2016;		

final product inspection result

Items	QC Standard	detection mode	result
1 Package	The packaging is intact without damage, and the batch number is correct.	visualization	<input checked="" type="checkbox"/> Pass
2 Appearance	The lyophilized powder was white in color, normal in morphology, and no collapse was observed.	visualization	<input checked="" type="checkbox"/> Pass
3 NTC	The sample solution was used as the template for testing, with the 6002R primer probe reaction buffer. No amplification was observed in the VIC channel, while the ROX channel showed normal amplification curves.	quantify PCR	<input checked="" type="checkbox"/> Pass

Quality inspection conclusion	This product was tested according to the enterprise standard, and the results comply with the regulations.		
QC Inspector	王晨光	QA Manager	陈普

