

Product Name	TOROIVD® Self-Collection Kit-For RNA-		
Cat NO	TPA-1611	QC Date	January 11th , 2023
Lot NO	R23010520	Report Date	January 11th , 2023
QC Items	End products	Store at	2°C-8°C for 2 years
According to	Enterprise standard ZS-B-ZLBZ-001	Expire Date	January 4th , 2025
QC Standard	According to the SOP process document developed by ISO13485: 2016;		

## QC Report

Items		QC Standard	Results
1	Package	The package is in good condition, no damage, no leakage, and the Lot number is correct.	<input checked="" type="checkbox"/> Pass
2	Appearance	Clear liquid, no precipitation, no suspended matter.	<input checked="" type="checkbox"/> Pass
3	pH	6.5 - 8	<input checked="" type="checkbox"/> Pass
4	Batch stability	The difference of Cq value of different batches of reagents at $\geq 50$ copies / reaction was not more than 0.5.	<input checked="" type="checkbox"/> Pass



QC Inspector		QA Manager	
--------------	--	------------	--