

Product Name		TOROIVD® Self-Collection Kit	
Cat NO	TPA-1620	QC Date	Jun. 28th , 2023
Lot NO	R010210	Report Date	Jun. 28th , 2023
QC Items	End products	Store at	Room temperature for 1 years
According to	Enterprise standard ZS-B-ZLBZ-001	Expire Date	Jun. 24th , 2024
QC Standard	According to the SOP process document developed by ISO13485: 2016;		

QC Report

Items		QC Standard	Results
1	Package	The package is in good condition, no damage, no leakage, and the Lot number is correct.	<input checked="" type="checkbox"/> Pass
2	Appearance	Clear liquid, no precipitation, no suspended matter.	<input checked="" type="checkbox"/> Pass
3	pH	7.5 - 8.5	<input checked="" type="checkbox"/> Pass
4	Batch stability	The difference of Cq value of different batches of reagents at ≥ 50 copies / reaction was not more than 0.5.	<input checked="" type="checkbox"/> Pass



QC Inspector		QA Manager	
--------------	--	------------	--